

miglionico

DENTAL EQUIPMENT



USER AND MAINTENANCE MANUAL



TABLE OF CONTENTS

1. DECLARATIONS OF CONFORMITY 4

2. SYMBOLS 6

3. INSTALLATION, TEST AND WARRANTY CERTIFICATE 7

 3.1 WARRANTY CONDITIONS 8

 3.2 INTENDED USE 8

 3.3 STANDARD AND OPTIONAL ACCESSORIES 8

 3.4 PRODUCT REGISTRATION AND CUSTOMER CARE 9

4. OPERATIONAL SETTING – ADVICES 11

 4.1 OPERATIONAL SETTING 11

 4.2 FIXING THE UNIT ONTO THE FLOOR 11

5. SUPPLY REQUIREMENTS AND INSTALLATION 12

6. TECHNICAL DATA 13

 6.1 UNIT 13

 6.2 CHAIR 13

7. CHAIR MOVEMENT LIMITS 13

8. DIMENSIONS 14

9. ESSENTIAL RECOMMENDATIONS FOR USE 15

10. IMPORTANT WARNINGS 15

11. PACKAGING DISPOSAL 16

12. DENTAL UNIT DISPOSAL 16

13. LABELS POSITION 17

 13.1 EXTERNAL LABELS 17

 13.2 FOOT CONTROL LABELS 18

 13.3 CHAIR BASE LABELS 18

 13.4 IDENTIFICATION LABELS 18

14. IDENTIFICATION LABEL 19

15. UNIT ON/OFF 19

16. INSTRUCTIONS FOR USE 20

17. NT CONTROL PANEL 20

 17.1 TOUCHSCREEN DISPLAY TURNING OFF 21

 17.2 LOCK SCREEN / CLEAN 21

 17.3 WI-FI FOOTPEDAL BATTERY LEVEL 21

 17.4 DATE AND TIME SETTINGS 21

 17.5 COUNTDOWN TIMER 21

 17.6 SETTING MENU 22

 17.6.1 TIMERS 22

 17.6.2 USERS 22

 17.7 CHAIR MOVEMENT AND MEMORIZING POSITIONS 23

 17.8 ADAPTIVE BACKREST POSITIONING 25

 17.9 BASIC FUNCTIONS 26

18. INSTRUMENTS 27

19. MCX BIEN AIR MICROMOTOR 27

 19.1 RESTORATIVE MODE 28

 19.2 ENDODONTICS MODE 29

20. MX2 BIEN AIR MICROMOTOR 31

 20.1 RESTORATIVE MODE 32

 20.2 ENDODONTICS MODE 33

 20.3 SURGERY MODE 35

20.4	PROGRAM MEMORIZATION.....	36
21.	SCALER.....	37
22.	ENDO SCALER (SATELEC).....	38
23.	TURBINE.....	39
24.	SYRINGE.....	40
25.	CURING LIGHT.....	40
26.	HANDPIECES.....	41
27.	SAFETY/EMERGENCY SYSTEMS.....	42
28.	HEADREST.....	43
29.	DENTIST’S CONSOLE CONFIGURATION.....	44
29.1	TOP DELIVERY VERSION.....	44
29.2	HANGING HOSES VERSION.....	45
29.3	DENTIST’S CONSOLE HANDLE REMOVAL.....	45
30.	PERISTALTIC PUMP.....	46
31.	ASSISTANT’S CONSOLE CONFIGURATION.....	47
31.1	ASSISTANT’S CONSOLE FUNCTIONS.....	47
32.	FOOT CONTROL FUNCTIONS.....	49
32.1	STANDARD FOOT CONTROL/ STANDARD WIRELESS FOOT CONTROL.....	49
32.2	PUSH FOOT CONTROL/ WIRELESS PUSH FOOT CONTROL.....	50
32.3	WIRELESS FOOT CONTROL INDICATIONS.....	51
33.	HOW TO CHOOSE THE INSTRUMENTS WATER SUPPLY SYSTEM.....	52
34.	MAINTENANCE.....	52
34.1	CLEANING AND DISINFECTION.....	52
34.2	CLEANING AND CARE OF THE UPHOLSTERY.....	53
34.3	STERILIZATION.....	54
35.	COMPULSORY TESTS.....	55
36.	SAFETY TEST REPORT.....	56
37.	COMPATIBILITY RATES.....	56
38.	TEST REPORTS AND WARNINGS.....	57
39.	MICROMOTOR RATIO VALUES.....	60
39.1	MCX MICROMOTOR.....	60
39.2	MX2 MICROMOTOR.....	61

1. DECLARATIONS OF CONFORMITY

DICHIARAZIONE DI CONFORMITA' UE

EU DECLARATION OF CONFORMITY

La Società MIGLIONICO S.R.L. (N. REG. Unico (SRN): in attesa di assegnazione), con sede legale e operativa Via Molise, Lotti 67/68 Z.I - 70021 Acquaviva delle Fonti (BA), dichiara, sotto la propria totale responsabilità, che il **dispositivo medico** denominato:

We undersigned MIGLIONICO S.R.L. (Single Registration Number (SRN): pending request), with head office addressed in Via Molise, Lotti 67/68 Z.I - 70021 Acquaviva delle Fonti (BA), declare under its own responsibility that the medical devices named:

	Nome commerciale / <i>Commercial Name</i>	Numero di serie / <i>Serial Number</i>	UDI-DI di base / <i>Basic UDI-DI</i>
Poltrona <i>Chair</i> (classe di rischio I) <i>(Risk class I)</i>	SYNCRO	S.N. XXXXX	805534993Z12110180D3

in accordo alla regola 13 dell'Allegato VIII, del Regolamento (UE) 2017/745 (MDR),

according to rule 13 of the Annex VIII of Regulation (EU) 2017/745 (MDR):

- è conforme ai requisiti essenziali ed alle disposizioni del Regolamento (UE) 2017/745 ess.mm.ii. come da Fascicolo Tecnico archiviato presso l'azienda;
comply with essential requirements and dispositions of Regulation (EU) 2017/745 (MDR) as per Technical File filed in company
- è fabbricato in accordo ai contenuti del Fascicolo Tecnico, che soddisfa i requisiti di cui all'Allegato II + III del sopra citato Regolamento.
is manufactured in compliance with the content of the Technical File, which satisfies the requirements of Annex II + III of the aforementioned Regulation.
- non sono state utilizzate Specifiche Comuni per la conformità dei suddetti dispositivi;
Common Specifications have not been used for the compliance of the aforementioned devices
- è conforme alla direttiva 2011/65/UE del Parlamento europeo e del Consiglio dell'8 giugno 2011, sulla restrizione dell'uso di determinate sostanze pericolose nelle apparecchiature elettriche ed elettroniche.
comply with Directive 2011/65 / EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment.

Acquaviva delle Fonti (BA), gg/mm/aaaa

Il legale rappresentante /
Legal Representative /

DICHIARAZIONE DI CONFORMITÀ UE

EU DECLARATION OF CONFORMITY

La Società Miglionico S.r.l. (codice SRN IT-MF-000019774), con sede legale ed operativa in Via Molise, Lotti 67/68 Z.I – 70021 Acquaviva delle Fonti (BA) in qualità di fabbricante dei **dispositivi Medici**:

We undersigned MIGLIONICO S.R.L., with head office addressed in Via Molise, Lotti 67/68 Z.I - 70021 Acquaviva delle Fonti (BA), as the manufacturer of the following medical devices named:

Descrizione generale	Nomi commerciali <i>Commercial Name</i>	Nomi commerciali <i>Commercial Name</i>	UDI-DI di Base/ <i>Basic UDI-DI</i>
Riuniti odontoiatrici "NICE"	NICE TOUCH NTX NICA TOUCH P19TP NICE TOUCH NTPX NICE TOUCH P19TPC NICE GLASS NGX NICE GLASS NGPX NICE GLASS P19G NICE GLASS P19GP NICE GLASS P19GPC NICE GLASSE F NICE TOUCH F NICE GLASS FP NICE TOUCH FP	NICE ONE NICE ONE P NICE ONE L NICE TOUCH NICE GLASS NICE TOUCH P NICE GLASS P NICE TOUCH P CART NICE GLASS P CART NICE TOUCH W NICE GLASS W NICE TOUCH P19T	805534993riunitinice9W

Destinati ad agevolare l'operatore per effettuare trattamenti nel cavo orale, di classe di rischio IIA, in accordo alla regola 12 dell'Allegato IX del Regolamento UE 2017/745, dichiara sotto la propria totale esclusiva responsabilità, che tali dispositivi:

Designed to facilitate the operator in treating the oral cavity, risk class IIA, according to rule 12 of annex IX To the EU Regulation 2017/745, declare under its own responsibility that the above-mentioned devices:

- sono conformi ai requisiti generali di sicurezza e prestazione ed alle disposizioni del Regolamento (UE) 2017/745 come da Documentazione Tecnica depositata presso l'ente e conservato presso la sede operativa del fabbricante; *comply with general requirements of safety and performance and other provisions of the EU Regulation 2017/745, as specified in the Technical File retained by the Notified body and held by the Company;*
- non sono state utilizzate Specifiche Comuni per la conformità dei dispositivi; *no Common Specifications have been used for the conformity of the devices;*
- sono fabbricati in accordo alla Documentazione Tecnica n. FT 001 MI, che soddisfa i requisiti di cui all'Allegato XI Parte A del sopra citato Regolamento, come da Certificato n. _____ rilasciato da ICIM S.p.a. con sede in Italia a Piazza Don Enrico Mapelli, 75 Sesto San Giovanni (MI) organismo Notificato 0425 il _____.
are manufactured according to the Technical File n. FT 001 MI, that complies with the provisions Mentioned in Annex XI Part A of the above-mentioned Regulation, as per Certificate No. _____, issued on _____ by the Notified Body No. 0425, ICIM SPA, with legal address: Piazza Don Enrico Mapelli, 75 Sesto San Giovanni, Italy.

Acquaviva delle Fonti (BA), gg/mm/aaaa

Il legale rappresentante /

Legal Representative /

2. SYMBOLS

	ALTERNATING CURRENT	Hz	MAINS FREQUENCY IN HERTZ
	GROUND PROTECTION	VA	MAXIMUM POWER ABSORBED BY THE UNIT (IN VA)
	B TYPE DEVICE	→	INDICATION
	ON/OFF	MDS	MIXER
	WARNING	AF	ANTI RETRACTION SYSTEM TO HANDPIECES
V	NOMINAL VALUE OF VOLTAGE (IN VOLTS)		DON'T LEAN OR PUSH

DISPLAY SYMBOLS

	SPRAY ON		X-RAY VIEWER		LOCK SCREEN / CLEAN
	SPRAY OFF		MICROMOTOR REVERSE ROTATION		INSTRUMENTS
	SPRAY ON/OFF		PERISTALTIC PUMP		ON/OFF OPTICAL FIBRE ON INSTRUMENTS
	COLD WATER CUP FILLER AND CUSPIDOR		UPWARD CHAIR MOVEMENT		PROPORTIONAL SPEED
	WARM WATER CUP FILLER AND CUSPIDOR		DOWNWARD CHAIR MOVEMENT		FIXED SPEED
	WATER TO CUSPIDOR		UPWARD BACKREST MOVEMENT		MEMORIZE POSITION
	WARM WATER CUP FILLER TIMER		DOWNWARD BACKREST MOVEMENT		OPERATING LIGHT
	COLD WATER CUP FILLER TIMER		ZERO POSITION		ASSISTANT CALL / DOOR OPENER
	WATER TO CUSPIDOR TIMER		EMERGENCY POSITION		COUNTDOWN TIMER
	TIMER DELAY CUSPIDOR AFTER CUP FILLER	PR	RINSING POSITION		SETTINGS
	OPTICAL FIBRE TIMER DELAY	RM	RECALL MEMORY		HOME

3. INSTALLATION, TEST AND WARRANTY CERTIFICATE

MIGLIONICO S.r.l.

sede legale e stabilimento
 via Molise, Lotti 67/68 Z.I. - 70021 Acquaviva delle Fonti (BA)
 Tel/fax 080759552 e-mail info@miglionico.net
www.miglionico.net

DICHIARA CHE I PRODOTTI

RIUNITO MODELLO <input type="checkbox"/> "NICE TOUCH" MATRICOLA _____ DATA COLLAUDO _____	POLTRONA MOD. "SYNCR0" MATRICOLA _____ DATA COLLAUDO _____
---	---

Sono stati installati seguendo le procedure standard della Miglionico srl

IL TECNICO timbro e firma _____	DATA INSTALLO _____
---	-------------------------------

DATI DEL CLIENTE Rag. Sociale _____ Via _____ Città _____ Tel/fax _____ e-mail _____	CONCESSIONARIO timbro e firma _____
--	---

- | | |
|--|--|
| Per ricevuta del manuale d'uso e manutenzione riunito | <input type="checkbox"/> SI <input type="checkbox"/> N |
| Per ricevuta del manuale d'uso e manutenzione telecamera | <input type="checkbox"/> SI <input type="checkbox"/> N |
| Per ricevuta del manuale d'uso e manutenzione lampada per compositi | <input type="checkbox"/> SI <input type="checkbox"/> N |
| Per ricevuta del manuale d'uso e manutenzione ablatore | <input type="checkbox"/> SI <input type="checkbox"/> N |
| Per ricevuta del manuale d'uso e manutenzione aspirazione chirurgica | <input type="checkbox"/> SI <input type="checkbox"/> N |

- Sottoscrivo e autorizzo Miglionico srl al trattamento dei miei dati personali nel rispetto della legge 675/96 e successive modificazioni del D.LGS 196/2003.
- Per accettazione delle condizioni di garanzia di seguito riportate.

CLIENTE Timbro e firma _____	L'AMMINISTRATORE UNICO MIGLIONICO SRL - timbro e firma _____
--	--

This document must be duly completed, signed and attached to the registration procedure accessible via QR code within 30 days from the installation date.

FAILURE TO INCLUDE THIS DECLARATION ON THE PLATFORM IMPLIES THE IMMEDIATE FORFEITURE OF THE THREE-YEAR WARRANTY EXTENDING.

3.1 WARRANTY CONDITIONS

The Dental unit comes with the "User and maintenance manual", "CE certification of compliance" and "installation, test and warranty certificate".

Warranty is valid for 12 months after installation date.

Customers who desire to extend their warranty must include the "installation, testing and warranty certificate", which must be duly completed, signed and attached to the registration procedure accessible via QR code within 30 days from the installation date. Failure to include the "installation, test and warranty certificate" in the registration form implies the immediate forfeiture of the right to the three-year extension of the warranty, as well as making it impossible for Miglionico s.r.l to comply with the legal obligations regarding the traceability of the medical device.

It remains confirmed that **the warranty covers all spare parts during the first year, including travel expenses and technical interventions, with the exception of material subject to wear and tear**, such as:

- hoses, aspiration cannulas and nozzles
- syringe tips
- scaler tips
- light bulbs, filters
- upholstery cuts
- painted parts damaged by bumps.

For the second, third and fourth year, the warranty is limited to spare parts produced by Miglionico, excluding handpieces and PCs, but the cost of the technical intervention including travel is borne by the customer.

The warranty is not restored with each replacement of spare parts.

Under no circumstances the replacement of the equipment or the supply of a replacement equipment is contemplated.

The 3-years warranty extension is invalid in case of:

- **"Installation, testing and warranty certificate" completed in its entirety not returned;**
- Repair or maintenance procedure carried out by unauthorized personnel;
- Accessories/spare parts not provided by the manufacturer have been implemented onto the unit;
- Damages caused by natural disaster, equipment misuse, negligence, incorrect installations, tampering, modification of the product, or the serial number, or accidental damage because of negligence of the client or third parties. Warranty also does not apply in the case of failures due to the electric supply more than indicated or sudden changes in electric voltage supply of the device connected, as well as in the case of failures caused by infiltration of liquids, fire, static discharge inductive / or electrostatic discharges caused by lightning, power surges or other external sources.

All relevant documentation related to handpieces, operating light, compressor, suction system etc. is considered as an integral part of this manual.

3.2 INTENDED USE

The equipment is destined to be used for diagnosis, prevention, check, therapy or cure of human disease of the oral cavity and oropharynx.

The device is used for different dental procedures, retraction of saliva, water, blood or other liquids used for the local treatment of the operated parts, scaling, cleaning, lighting coverage of the oral cavity.


3.3 STANDARD AND OPTIONAL ACCESSORIES

When placing the purchase order, each dental unit is configured according to the customer's request and is equipped with the required accessories, divided into the following categories:

- DENTIST CONSOLE CONFIGURATION
- ASSISTANT CONSOLE CONFIGURATION
- HYDRO CASE CONFIGURATION

3.4 PRODUCT REGISTRATION AND CUSTOMER CARE

To better manage the feedback and requests of end customers and to guarantee the traceability of products thanks to the acquisition of data relating to installations, Miglionico has finally established its Customer Care Department. The contact details of the new office are as follows.



Maria Mastrorocco
 customercare@miglionico.net
 +39 3427835677

Furthermore, to facilitate the product registration operation which allows the customer to access the free 3-year warranty extension in addition to the standard 1-year warranty period, starting from unit s.n.: 24SY-04692 (production week n.17-2024), you will find the following card on the dentist’s console of the unit instead of the extension request form:



GRAZIE PER AVER SCELTO IL NOSTRO PRODOTTO • **THANK YOU FOR CHOOSING OUR PRODUCT**
MERCI D'AVOIR CHOISI NOTRE PRODUIT • **GRACIAS POR ELEGIR NUESTRO PRODUCTO**
VIELEN DANK, DASS SIE SICH FÜR UNSER PRODUKT ENTSCHEIDEN HABEN

Per beneficiare dell'estensione di garanzia registra entro 2 mesi il tuo prodotto
 Register your product within 2 months to obtain the free warranty extension

REGISTRA IL TUO RIUNITO • **REGISTER YOUR UNIT**
 Enregistrez votre produit dans un délai de 2 mois pour bénéficier de l'extension de garantie gratuite
 Registra tu producto dentro de 2 meses para beneficiar de la extensión de garantía gratuita

ENREGISTREZ VOTRE UNITÉ • **REGISTRA TU EQUIPO**
 Registrieren Sie Ihr Produkt innerhalb von 2 Monaten, um von der Garantieverlängerung zu profitieren

REGISTRIEREN SIE IHRE BEHANDLUNGSEINHEIT

SCANSIONA IL QR CODE SUL RETRO • SCAN THE QR CODE ON THE BACK
 SCANNER LE QR CODE AU DOS • ESCANEA EL CÓDIGO QR EN LA PARTE TRASERA
 SCANNEN SIE DEN QR-CODE AUF DER RÜCKSEITE



REGISTRA IL PRODOTTO
 PER OTTENERE L'ESTENSIONE DI GARANZIA GRATUITA
REGISTER YOUR PRODUCT
 TO OBTAIN THE FREE WARRANTY EXTENSION
ENREGISTREZ VOTRE PRODUIT
 POUR OBTENIR L'EXTENSION DE GARANTIE GRATUITE
REGISTRA TU PRODUCTO
 PARA OBTENER LA EXTENSION DE GARANTÍA GRATUITA
REGISTRIEREN SIE DAS PRODUKT,
 UM DIE KOSTENLOSE GARANTIEVERLÄNGERUNG ZU ERHALTEN

SCARICA IL MANUALE D'USO
DOWNLOAD THE USER MANUAL
TÉLÉCHARGEZ
LE MANUEL D'UTILISATION
DESCARGA EL MANUAL DE USUARIO
LADEN SIE DAS
HANDBUCH HERUNTER

SODDISFATTO DELL'ACQUISTO?
 LASCIACI LA TUA OPINIONE
ARE YOU HAPPY WITH YOUR PURCHASE?
 LEAVE US YOUR OPINION
ÊTES-VOUS SATISFAITS DE VOTRE ACHAT?
 LAISSEZ VOTRE ÉVALUATION
¿ESTÁS SATISFECHO DE TU COMPRA?
 DEJA TU OPINIÓN
SIND SIE MIT DEINER ANSCHAFUNG ZUFRIEDEN?
 HINTERLASSEN SIE IHRE MEINUNG

To register the product, the customer must follow the following steps:

- Scan the first QR code with the serial number with your smartphone camera;
- Fill in all the fields with your clinic's information;
- Enter the date of installation and attach a photo of the INSTALLATION, VERIFICATION AND WARRANTY FORM filled out by the technician who installed the unit;
- Accept the privacy conditions and click on "send".

The warranty extension certificate will be sent to the email indicated during registration. It will no longer be necessary to send the installation form via email to obtain the extension.

NOTE: don't forget the other QR codes on the card! Remind your customers to download the user manual via the central QR code and to leave us a review if they are satisfied with their purchase! ★★ ★★ ★★ ★★ ★★

Furthermore, the following QR code will be present inside the hydro case to allow the customer to consult the user manual at any time.



Finally, in order to detect customer satisfaction with the new production line, the Customer Care will take care of calling your customers for a customer satisfaction survey regarding the products they purchased.

For further information or particular needs, you can contact the Sales Director Katia Cea at k.cea@miglionico.net or by phone (+39 3349910634) or our Customer Care Dept. customercare@miglionico.net

4. OPERATIONAL SETTING – ADVICES

4.1 OPERATIONAL SETTING


The operating spaces must respect the regulations about the application of the product: minimal dimension of the rooms mq 7,5; long side Mt 3

Washable floor, advised illumination with fluorescent tubes 5500°k.

The systems (electrical, water-sewer, compressed air and surgical suction) have to be realized according to the regulations in force.

4.2 FIXING THE UNIT ONTO THE FLOOR

Warning: it is recommended to fix the unit onto the floor by using 2pcs 10 mm fisher screws



5. SUPPLY REQUIREMENTS AND INSTALLATION

OBJECT	SPECIFIC DESCRIPTIONS						
Location	Relative humidity between 45% and 75% Temperature between 15 °C and 35 °C Air pressure between 860mbar÷ 1060 mbar (645 mmHg to 795 mmHg)						
Water supply 	Water has to be adequate to the national rules for the drinking water. The water supply shall use drinking water, filtered and decalcified, for domestic use. The water shall have the following features: <ul style="list-style-type: none"> • Hardness between 15 ÷ 20 F° (French degree) • Pressure between 150 ÷ 400 kPa (1.5 ÷ 4 bar) • Capability > 3l/min to 400 kPa (4 bar) In case of pressures higher than 400 kPa (4 bar) they have to insert an adequate pressure reducer before the unit installation. The supply tube has to be equipped with an arrest tap. Before installation technicians must clean the tubes with care in order to avoid the possible penetration of impurities into the hydro case of the unit by purging it until the elimination of the impurities.						
Electric system 	Complies with current regulations (regulations for electrical systems in rooms used for medical purposes for type 'A' medical clinics) on the date of installation. The single-phase mains voltage 230V ± 10% frequency 50Hz.						
Electric supply 	As indicated in the device data tag. Allowed tolerance on ± 10% supply voltage. Absorbed power full-load 1400 VA The dental unit is equipped with a proper supply terminal board for a permanent connection to the power network, which has to have a 10 A - 250 V bipolar switch with differential intervention current IΔN=0.03 A, made in accordance with the European regulations concerning the device.						
Suction system	Ejection air tube of the suction system should discharge air outside the housing rooms, for hygiene and environmental reasons. Nether or external tube ought to have a 350 l/min air capacity and a 20kPa (0.2 bar) low pressure value.						
Air supply	The compressor must be placed in an open room, hygienically and protected from heat sources to not pick up air discharge from the surgical aspirator. Air pressure ought to be included between 500kPa and 700 kPa (5 ÷ 7 bar). Major or equal 60l/min a 500 kPa (5 bar) capacity. Compressor equipped with air dry system and antibacterial filter. Supply tube must have an arrest tap.						
Drain pipes	Piping has to be realized with a PVC (or higher quality) tube. Piping must have a slope not less than 1.5 cm each meter and a siphon that allows the inspection every 4 metres if the distance to the upright column is higher.						
Weights and dimensions	<table style="width: 100%; border: none;"> <tr> <td style="width: 60%;">Chair weight</td> <td style="text-align: right;">Kg. 126</td> </tr> <tr> <td>Chair + dental unit weight</td> <td style="text-align: right;">Kg. 197</td> </tr> <tr> <td>Total weight (Chair, unit, operating light)</td> <td style="text-align: right;">Kg. 204</td> </tr> </table>	Chair weight	Kg. 126	Chair + dental unit weight	Kg. 197	Total weight (Chair, unit, operating light)	Kg. 204
Chair weight	Kg. 126						
Chair + dental unit weight	Kg. 197						
Total weight (Chair, unit, operating light)	Kg. 204						

6. TECHNICAL DATA

6.1 UNIT

The equipment is not proper to be used in presence of anaesthetic inflammable mixture with oxygen air or nitrous oxide.


MODELLO	NICE GLASS (NGF) – NICE TOUCH (NTF)
CLASSIFICAZIONE (EN 60601-1)	Classe I Tipo B 
CLASSIFICAZIONE (93/42 CEE)	Classe II a
TENSIONE DI ALIMENTAZIONE	230 V
CORRENTE ALTERNATA MONOFASE	50/60 Hz
POTENZA ASSORBITA A PIENO CARICO	1400 VA
PESO ADDIZIONALE SOPPORTATO DAL VASSOIO PORTA-TRAY	KG 1,5

6.2 CHAIR

The equipment is not proper to be used in presence of anaesthetic inflammable mixture with oxygen air or nitrous oxide.

MODEL	SYNCRO (NSY)
CLASSIFICATION (EN 60601-1)	Class I Type B 
CLASSIFICATION (93/42 CEE)	Class I
SUPPLY VOLTAGE	230 V
SINGLE PHASE ALTERNATIVE CURRENT	50 Hz
INTERMITTENT FUNCTIONING	18 min. of rest every 3 min. working
CHAIR MINIMUM HEIGHT	410 mm
CHAIR MAXIMUM HEIGHT	900 mm
BACKREST RISING MOTOR	ELECTRIC MOTOR 32 Vdc MAX 10,5 A
BACKREST MOTOR	ELECTRIC MOTOR 32 Vdc MAX 5,2 A
MAXIMUM LOAD CAPACITY	KG 180
PROTECTION LEVEL AGAINST LIQUIDS PENETRATION	IPX0 – NO PROTECTION


7. CHAIR MOVEMENT LIMITS



CHAIR MOVEMENT UP AND DOWN
min. 410mm
max. 900mm

8. DIMENSIONS

IT - DIMENSIONI D'INGOMBRO
 ENG - OVERALL DIMENSIONS
 FRA - DIMENSIONS D'ENCOMBREMENT
 ESP - DIMENSIONES GLOBALES
 DE - GESAMTGRÖßE



misure in millimetri
 measurements in millimeters
 mesures en millimètres
 medidas en milímetros
 maße in millimeter

9. ESSENTIAL RECOMMENDATIONS FOR USE

This equipment is corresponded to the quality standard CEI EN 60601-1 (general standards for electro - medical devices safety) and CEI EN 60601-1-2 the standards corresponded to the CE Directive 93/42, this equipment is exclusively destined to the dentist use, potentially assisted by auxiliary personnel, prepared for dental assistance.

It is necessary to study the user manual and to read carefully all the instructions concerning the instruments.

After the unit is installed but before using it, it's necessary to follow the instructions down below:

- Sterilize operational instruments, which are not in sterile packaging (ref. to the “operational tools” section)
- Sterilize at 135°C in the autoclave the removable parts in silicon (handles and carpet)
- Disinfect all the parts normally are not come into contact with the patient (ref. to the “cleaning and disinfection” section)
- Activate the water to the glass and to the handpieces with spray at least for 3 minutes, so the disinfection liquid starts to circulate.
- Remove all handpieces and dental cutter after every operation.

Protect eyes, respiratory tracts, mouth and skin by wearing glasses, special mask and disposable gloves to protect from fragments coming from the patient mouth. Moreover, use the aspirator at high speed in order to suck the dust and the little particles released in the air during the handpieces use.

The patient should not wear short clothes to preserve the hygienic conditions.

WARNING: Remove handpieces (micromotor handpieces, turbine handpieces, scaler handpiece, optic terminal of the lamp, syringe covers, silicon handles and carpet under the instruments) from their lodging after every operation in order to realize their sterilization and avoid cross-contaminations.



10. IMPORTANT WARNINGS

Miglionico s.r.l is not to be held liable for safety, reliability and performances if:

- the installation operations have not been carried out by qualified personnel authorized by MIGLIONICO and equipped with license.
- the electric, water, air supply systems, the water discharge system, the possible suction system and the rooms where the device is installed do not comply with the law.
- non-authorized modifications have been carried out (including connecting other medical devices or accessories) and/or non-original spare parts have been used.
- the device is not used as it is recommended in the user and maintenance manual.
- the annual technical maintenance is not carried out respecting the set date by technicians.
- **the water supply system is left open without any personnel being present.**

Before activating the chair, please make sure that there are no any other equipment or accessories which can obstruct the chair movement.

WARNING: do not remove any carter before switching off the dental unit.



WARNING: the front carter of the base can be removed **ONLY** by authorized technicians because, even if the dental unit is switched off, there are some elements under voltage and there is an electric shock danger.

For the use, maintenance, sterilization and cleaning of the handpieces, please read the instructions in their packaging. Miglionico s.r.l is not to be held responsible for possible damages that caused by the inobservance and by the omission of the above-mentioned rules.



11. PACKAGING DISPOSAL


The materials used for the package are 100% recyclable and they have to be brought to an authorized garbage dump which will provide for the recycle or the disposal.

12. DENTAL UNIT DISPOSAL

When the dental unit is permanently off duty, please remove the supply cables and the fuses and break the electric parts in an irreparable way before delivering it to a garbage dump authorized for the recycle of materials.

13. LABELS POSITION

13.1 EXTERNAL LABELS




1	MIGLIONICO LOGOTYPE	
2	SAFETY DEVICE	
3	CE MARK	
4	NTF MODEL LOGO	
5	ON/OFF	
6	IDENTIFICATION LABEL	
7	USER MANUAL QR CODE	

13.2 FOOT CONTROL LABELS




1	RECALL MEMORY: PATIENT DESCENT WATER TO HANDPIECES ON/OFF (WHEN THE INSTRUMENT IS SELECTED)	
2	RINSING POSITION	
3	JOYSTICK FOR CHAIR MOVEMENTS	

13.3 CHAIR BASE LABELS



1	GENERAL SUPPLY INPUT	
---	----------------------	---

13.4 IDENTIFICATION LABELS



1	IDENTIFICATION TAG OF THE UNIT	
---	--------------------------------	---

14. IDENTIFICATION LABEL

The dental unit is traceable by a serial number printed on the label. Please include your serial number when requesting information or spare parts.

Read par. 13.4 and 13.1 (5).




15. UNIT ON/OFF

The unit is equipped with a main switch located in the lower part of the chair. When you push and release the button, the button will turn green and two segments on the control panel display will light up, indicating the electricity supply activation. By pushing and releasing the general switch again, the unit will be turned off.

It is recommended to turn the main switch off every time you stop working or in any case, before every technical or maintenance operations that imply the intervention on the parts protected by carters.

For what concerns the NT models, you must turn off the unit by the dedicated function displayed on the control panel before turning of the main switch.



WARNING: The front carter at the base of the chair can be removed by authorized technician only because, even if the dental unit is switch off, there are some elements under voltage and there is an electric shock danger.



16. INSTRUCTIONS FOR USE

Operational instruments need to be withdrawn from their initial position and activated the foot control (see foot-control functions). The air-water syringe is not activated by the pedal but directly by using the buttons on it.


The dental unit is equipped with a functioning instruments priority system. Only the first selected instrument is active, and the rest of the instruments are blocked.

The **anti-retraction system (AF)** reduces to the minimum the concentration of liquids or debris coming from the operating area to the spray holes. This consists in the release of a jet of air every time the pedal lever is released, with the function “instrument spray on” or “instrument spray on/off” is active.

ATTENTION: When an instrument is removed from its position and is activated by the pedal, the chair movement controls are deactivated. This is to avoid accidental movements of the chair while operating on the patient.



17. NT CONTROL PANEL



17.1 TOUCHSCREEN DISPLAY TURNING OFF

	By pressing the icon for 5 seconds, the touch panel turns off. After that, it is possible to turn off the dental unit from the main switch.
---	---

17.2 LOCK SCREEN / CLEAN

	By pressing the icon for 3 seconds, a 20-second screen lock is activated, which allows cleaning the display easily.
---	---

17.3 WI-FI FOOTPEDAL BATTERY LEVEL

	WIRELESS foot pedal battery charge level indicator. It is green if the battery is charged and turns red when it is time to recharge the pedal. Present exclusively with WIRELESS pedal.
---	---


17.4 DATE AND TIME SETTINGS

On the display, in any selected screen, the top bar with DATE and TIME will always be visible. To set the date and time, click on the bar, and a screen with a calendar and time setting will appear. Select the desired date and time and click on the icon at the bottom right corner to save. If you accidentally press on the date and time bar, just click on the save icon without making any changes or press on the bar again.



17.5 COUNTDOWN TIMER


Thanks to the COUNTDOWN section, it is possible to set a timer that starts a countdown. When tools are removed (for example the curing light), the countdown timer will continue going down and will be visible on the upper bar in place of the timer icon. At the end of the countdown, an alarm sound will go off.



17.6 SETTING MENU

Thanks to the SETTINGS folder, it is possible to set the timers relating to the water to cup filler, cuspidor water and LED light for the instruments, and it is also possible to manage the users.

17.6.1 TIMERS





	WARM WATER CUP FILLER TIMER It allows you to adjust the time for dispensing hot water to the cup. By using the cursor, you can set a time ranging from 1 to 15 seconds. NOT USED
	COLD WATER CUP FILLER TIMER It allows you to adjust the time for dispensing cold water to the cup. By using the cursor, you can set a time ranging from 1 to 15 seconds. NOT USED
	CUSPIDOR TIMER It allows you to adjust the time for dispensing water to the cuspidor. By using the cursor, you can set a time ranging from 1 to 15 seconds. NOT USED
	OPTICAL FIBER TIMER It allows you to adjust the delay time in turning off the LED light on the instruments. By using the cursor, you can set a time ranging from 1 to 15 seconds.
	TIMER DELAY CUSPIDOR AFTER CUP FILLER It allows you to set a time, up to 15 seconds, to delay the flushing water to the cuspidor after filling the cup. By moving the cursor towards the minimum, the icon will indicate "OFF", therefore water will be flushed to the cuspidor immediately after the cup is filled. NOT USED

17.6.2 USERS

The unit can manage up to 4 users. Personalized parameters will be stored for each user, such as chair positions (1, 2, 3, 4) or micromotor programs.


Furthermore, you have the possibility to rename users by long-pressing on the selected user.




17.7 CHAIR MOVEMENT AND MEMORIZING POSITIONS

In every instrument section, it is possible to manage chair movements and activate the basic functions of the unit by using the icons above and below the home and instrument sections.

In the HOME screen, it is possible to move the chair using the arrows corresponding to the chair parts and recall the memorized positions by the icons in the top bar. In the instrument sections, it is possible to activate the same functions thanks to the scroll bar. If the instrument is active, chair movements are blocked. If the instrument is picked up but is not active, movements can be activated.



		UPWARD CHAIR MOVEMENT It rises the height of the seat. When a memory that includes this movement is recalled, this icon is activated.
		DOWNWARD CHAIR MOVEMENT It lowers the height of the seat. When a memory that includes this movement is recalled, this icon is activated.
		DOWNWARD BACKREST MOVEMENT It reclines the backrest. When a memory that includes this movement is recalled, this icon is activated.
		UPWARD BACKREST MOVEMENT It rises the backrest towards the seat. When a memory that includes this movement is recalled, this icon is activated.



	<p>MEMORIZATION</p> <p>Thanks to this icon, it is possible to memorize the 4 positions of the chair and customize the positions already memorized by the system: EMERGENCY position, RESET position and RINSE position. Move the chair to the desired position, press the MEMORIZE POSITION icon, and press the icon in which you want to store the position within 3 seconds (1, 2, 3, 4 or position icons).</p>
	<p>WARNING: when memorizing a position, never bring the seat and backrest to their maximum limit; adjust to a few millimetres from the maximum limit and then memorize.</p>
RM	<p>RECALL MEMORY</p> <p>It recalls the 4 previously saved working positions. Press the RM icon and then the position you want to recall within 3 seconds. If you do not press one of the positions after 3 seconds, the icons will automatically disappear.</p>
PR	<p>RINSING POSITION</p> <p>It moves the backrest to the rinsing or patient rest position. A subsequent press on the icon brings the backrest back to the zero position. If after activating the position you do not want to return to the previous one, you must press one of the 4 chair movement buttons to reset the function.</p> <p>You can customize this position thanks to the MEMORIZE POSITION icon. Move the backrest to the most comfortable position for the patient when he needs to rinse and press PR within 3 seconds. In this way, the position will be memorized and activated every time this icon is pressed.</p>
	<p>RESET/ZERO POSITION</p> <p>It brings the chair to the zero position, to help the patient sit or get up from it. You can customize this position thanks to the MEMORIZE POSITION icon. Move the chair to the most comfortable position for the patient to sit/get up and press RESET/ZERO POSITION within 3 seconds. In this way, the position will be memorized and activated every time this icon is pressed.</p>
	<p>EMERGENCY POSITION</p> <p>It brings the chair to the Trendelenburg position. You can customize this position thanks to the MEMORIZE POSITION icon. Move the backrest to the desired position and press EMERGENCY within 3 seconds. In this way, the position will be memorized and activated every time this icon is pressed.</p>

17.8 ADAPTIVE BACKREST POSITIONING


To use the adaptive backrest movement function in one of the three available positions, the chair must be empty and the patient must not be sitting in it. Then, click on the backrest icon at the bottom left of the display (Fig. A) and select the desired height among:

- **Short** (suitable for children and people with below average height);
- **Medium** (suitable for people with an average height);
- **Tall** (suitable for particularly tall people).

Once the desired height has been chosen, the backrest automatically positions itself in the ideal position and the chosen backrest option will be shown on the display. (Fig. B)




17.9 BASIC FUNCTIONS



	<p>ASSISTANT CALL / DOOR OPENER It is possible to associate this command to an external low voltage device, which can be, for example, a sound device placed in another room or an automatic door opening device.</p>
	<p>X-RAY VIEWER It activates/disactivates the X-ray film viewer screen.</p>
	<p>WATER TO CUSPIDOR Activate the water flush to the cuspidor for a set time (to set the cuspidor timer, see par. 17.6.1 TIMER). NOT USED</p>
	<p>WARM WATER CUP FILLER AND CUSPIDOR It fills the cup with warm water and then activates the cuspidor flushing for a set time and after a set time. (for glass and cuspidor timer settings, see par. 17.6.1 TIMER). Pressing the icon again will stop the flushing. NOT USED</p>
	<p>COLD WATER CUP FILLER AND CUSPIDOR It fills the cup with cold water and then activates the cuspidor flushing for a set time and after a set time. (for glass and cuspidor timer settings, see par. 17.6.1 TIMER). Pressing the icon again will stop the flushing. NOT USED</p>
	<p>OPERATING LIGHT It turns the operating light on/off. If it's off, when the memorized positions (1, 2, 3, 4) are recalled, it will turn on automatically. If it's on and you activate the RINSING or RESET positions, it will turn off at the end of the movement. NOT USED</p>



18. INSTRUMENTS

All the functions and parameters relating to the instruments can be set and managed by the screen which automatically appears when withdrawing the instrument or by the INSTRUMENTS section when the instrument is in its housing on the dentist console. This section allows you to access the screens of all instruments of the dentist console and change their settings without having to withdraw the instrument. The parameters saved in the instruments section activate automatically when the instrument is used. The instrument screens are in the same order as the instruments are positioned on the console. To switch from one instrument to another, swipe right or left.




19. MCX BIEN AIR MICROMOTOR

When you withdraw it from its instrument housing, it is activated and regulated by the foot pedal lever. It can be used in two different modes: Restorative and Endodontics, which can be selected from the drop-down menu at the top right.



19.1 RESTORATIVE MODE

The Restorative mode allows you to adjust the speed of the micromotor, the type of handpiece used, manage the peristaltic pump, the spray, the optical fibre and the inversion of the rotating direction of the micromotor. By activating the pedal lever, the micromotor turns progressively from minimum rpm to the set value.



FUNCTIONS


	<p>INSTRUMENT. It identifies which instrument you are using. If there are several identical instruments, the indicator will mark 1 / 2 / 3 according to their position on the console.</p>
	<p>TYPE OF HANDPIECE. By pressing this icon, it is possible to select which type of reducing or multiplying handpiece is being used on the micromotor. As the ratio changes, the number of rpms will vary. (For the list of handpieces and the related rpms of the micromotor, see par. 46)</p>
	<p>REAL MICROMOTOR RPM. By activating the pedal lever, the micromotor rotates progressively from the minimum speed to the set value. The rpms can be adjusted by the cursor, the + and - keys and the shortcut keys. The micromotor speed limit depends on the ratio of the handpiece used (for the rpm values associated with the handpiece ratio, see par. 46).</p>
	<p>PERISTALTIC PUMP. If it's not included on the unit, the icon will be deactivated. If it's present, the icon will turn the peristaltic pump on/off. When you activate it, a bar will appear to let you adjust the percentage of distilled water flow. The instrument spray must also be active for correct irrigation.</p>
	<p>INSTRUMENT SPRAY. SPRAY ON: Air-water spray active when the pedal lever moves to the right. By pressing it again, you will activate the SPRAY ON/OFF function: the air-water spray will activate when you move the pedal lever to the right and press the steel lever down at the same time. By pressing it again, you will deactivate the function: SPRAY OFF.</p>
	<p>OPTICAL FIBRE. It turns on / off the LED light on the handpiece if the OPTICAL FIBRE is present on the instruments. To adjust the LED light delay, see par. 17.6.1 TIMER.</p>
	<p>MICROMOTOR ROTATION INVERSION. This function inverts the rpms rotating direction of the micromotor.</p>

19.2 ENDODONTICS MODE

The Endodontics mode allows you to adjust the speed of the micromotor, the Torque value, the type of handpiece used, activate the auto reverse, auto stop and auto forward functions with related reverse time and cruise control, manage the peristaltic pump, spray, the optical fibre and inverse the micromotor rotating direction.

All these parameters can be memorized in 10 customizable and recallable programs.

By activating the pedal lever, regardless of its position, the micromotor turns at the set speed.




FUNCTIONS

<p>MCX - 1</p>	<p>INSTRUMENT. It identifies which instrument you are using. If there are several identical instruments, the indicator will mark 1 / 2 / 3 according to their position on the console.</p>
<p>1:1</p>	<p>TYPE OF HANDPIECE. By pressing this icon, it is possible to choose which type of reducing or multiplying handpiece is being used on the micromotor. As the ratio changes, the number of rpms and the micromotor torque limits will vary. (For specific speed and torque values of handpieces, see par. 46)</p>
<p>PROGRAM 01</p>	<p>PROGRAM MENU. It indicates the selected program. (see par. 20.4)</p>
<p>AUTO FORWARD</p> <p>5s</p>	<p>AUTO FORWARD. When you move the foot pedal lever to the right, the micromotor turns clockwise, whereas it inverts the rotating direction when it reaches the selected torque at the set time via the slide bar and then turns clockwise again. Only this function can activate the CRUISE CONTROL function.</p>
<p>AUTO REVERSE</p>	<p>AUTO REVERSE. When you move the foot pedal lever to the right, the micromotor turns clockwise, whereas it inverts the rotating direction when it reaches the set torque until you release the foot pedal.</p>
<p>AUTO STOP</p>	<p>AUTO STOP. When you move the foot pedal lever to the right, the micromotor turns clockwise and stops when it reaches the set torque. To restart the rotation, you must set the pedal lever to zero and bring it back to the right.</p>

	<p>MICROMOTOR RPM. By activating the foot pedal lever, regardless of its position, the micromotor turns at the set speed. The speed can be adjusted by the cursor, the + and – keys and the hotkeys. The speed limit of the micromotor depends on the ratio of the handpiece used. (for the speed values associated with the handpiece ratio, see par. 46)</p>
	<p>TORQUE VALUE. It's a settable value ranging from 0.25Ncm to 53Ncm, depending on the handpiece used (for the torque values associated with the handpiece ratio, see par. 46). When the set resistance is reached, the micromotor will react according to the type of function activated (AUTO FORWARD, AUTO STOP or AUTO REVERSE).</p>
	<p>PERISTALTIC PUMP. If it's not included on the unit, the icon will be deactivated. If it's present, the icon will turn the peristaltic pump on/off. When you activate it, a bar will appear to let you adjust the percentage of distilled water flow. The instrument spray must also be active for correct irrigation.</p>
	<p>CRUISE CONTROL. It can be exclusively activated in ENDODONTICS mode and with AUTO FORWARD function active. This function allows the micromotor to continue turning even if you release the foot pedal lever after activating it by pushing the foot pedal lever to the right.</p> <ul style="list-style-type: none"> - Activate the cruise control function by pressing the icon; - Once the icon is pressed, an alert will pop up. To make sure that the activation of this function is desired, press YES; - The activated icon will flash red; - Move the lever to the right to activate the micromotor; - The micromotor will continue to turn even if you release the foot pedal; - To stop the micromotor while this function is activated, move the pedal lever to the right; - To deactivate this setting, press the icon again.
	<p>ATTENTION: pay particular attention to the micromotor when using it with this function active, remember that the micromotor will continue to turn if this function is not deactivated. It could be dangerous for the operator and the patient if not used with care.</p>
	<p>INSTRUMENT SPRAY. SPRAY ON: Air-water spray active when the pedal lever moves to the right. By pressing it again, you will activate the SPRAY ON/OFF function: the air-water spray will activate when you move the pedal lever to the right and press the steel lever down at the same time. By pressing it again, you will deactivate the function: SPRAY OFF.</p>
	<p>OPTICAL FIBRE. It turns on / off the LED light on the handpiece if the OPTICAL FIBRE is present on the instruments. To adjust the LED light delay, see par. 17.6.1 TIMER.</p>
	<p>MICROMOTOR ROTATION INVERSION. This function inverts the rmps rotating direction of the micromotor.</p>


20. MX2 BIEN AIR MICROMOTOR

When you withdraw it from its instrument housing, it is activated and regulated by the pedal lever. It can be used in three different modes: Restorative, Endodontics and Surgery. These functions can be selected from the drop-down menu at the top right corner.



20.1 RESTORATIVE MODE

The Restorative mode allows you to adjust the speed of the micromotor, the type of handpiece used, manage the peristaltic pump, the spray, the optical fibre and the inversion of the rotating direction of the micromotor. By activating the pedal lever, the micromotor rotates progressively from minimum rpm to the set value.



FUNCTIONS


	<p>INSTRUMENT. It identifies which instrument you are using. If there are several identical instruments, the indicator will mark 1 / 2 / 3 according to their position on the console.</p>
	<p>TYPE OF HANDPIECE. By pressing this icon, it is possible to choose which type of reducing or multiplying handpiece is being used on the micromotor. As the ratio changes, the number of rpms and the micromotor torque limits will vary. (For specific speed and torque values of handpieces, see par. 46)</p>
	<p>MICROMOTOR RPM. By activating the foot pedal lever, regardless of its position, the micromotor turns at the set speed. The speed can be adjusted by the cursor, the + and – keys and the hotkeys. The speed limit of the micromotor depends on the ratio of the handpiece used. (for the speed values associated with the handpiece ratio, see par. 46).</p>
	<p>PERISTALTIC PUMP. If it's not included on the unit, the icon will be deactivated. If it's present, the icon will turn the peristaltic pump on/off. When you activate it, a bar will appear to let you adjust the percentage of distilled water flow. The instrument spray must also be active for correct irrigation.</p>
	<p>INSTRUMENT SPRAY. SPRAY ON: Air-water spray active when the pedal lever moves to the right. By pressing it again, you will activate the SPRAY ON/OFF function: the air-water spray will activate when you move the pedal lever to the right and press the steel lever down at the same time. By pressing it again, you will deactivate the function: SPRAY OFF.</p>
	<p>OPTICAL FIBRE. It turns on / off the LED light on the handpiece if the OPTICAL FIBRE is present on the instruments. To adjust the LED light delay, see par. 17.6.1 TIMER.</p>
	<p>MICROMOTOR ROTATION INVERSION. This function inverts the rpms rotating direction of the micromotor.</p>

20.2 ENDODONTICS MODE

The Endodontics mode allows you to adjust the speed of the micromotor, the Torque value, the type of handpiece used, activate the auto reverse, auto stop and auto forward functions with related reverse time and cruise control, manage the peristaltic pump, spray, the optical fibre and inverse the micromotor rotating direction.

All these parameters can be memorized in 10 customizable and recallable programs.

By activating the pedal lever, regardless of its position, the micromotor rotates at the set speed.



FUNCTIONS

<p>MX2 - 1</p>	<p>INSTRUMENT. It identifies which instrument you are using. If there are several identical instruments, the indicator will mark 1 / 2 / 3 according to their position on the console.</p>
<p>1:1 ▾</p>	<p>TYPE OF HANDPIECE. By pressing this icon, it is possible to choose which type of reducing or multiplying handpiece is being used on the micromotor. As the ratio changes, the number of rpms and the micromotor torque limits will vary. (For specific speed and torque values of handpieces, see par. 46)</p>
<p>PROGRAM 01 ▾</p>	<p>PROGRAM MENU. It indicates the selected program. (see par. 20.4)</p>
<p>AUTO FORWARD ▾ 5s</p>	<p>AUTO FORWARD. When you move the foot pedal lever to the right, the micromotor turns clockwise, whereas it inverts the rotating direction when it reaches the selected torque at the set time via the slide bar and then turns clockwise again. Only this function can activate the CRUISE CONTROL function.</p>
<p>AUTO REVERSE ▾</p>	<p>AUTO REVERSE. When you move the foot pedal lever to the right, the micromotor turns clockwise, whereas it inverts the rotating direction when it reaches the set torque until you release the foot pedal.</p>
<p>AUTO STOP ▾</p>	<p>AUTO STOP. When you move the foot pedal lever to the right, the micromotor turns clockwise and stops when it reaches the set torque. To restart the rotation, you must set the pedal lever to zero and bring it back to the right.</p>
<p>RECIPROCAL ▾</p>	<p>RECIPROCAL. This function will make the motor rotate with alternating rotation movements, with speed and torque pre-set automatically. When this mode is active, it is not possible to make changes to the other parameters.</p>

	<p>MICROMOTOR RPM. By activating the foot pedal lever, regardless of its position, the micromotor turns at the set speed. The speed can be adjusted by the cursor, the + and – keys and the hotkeys. The speed limit of the micromotor depends on the ratio of the handpiece used. (for the speed values associated with the handpiece ratio, see par. 46)</p>
	<p>TORQUE VALUE. It's a settable value ranging from 0.25Ncm to 53Ncm, depending on the handpiece used (for the torque values associated with the handpiece ratio, see par. 46). When the set resistance is reached, the micromotor will react according to the type of function activated (AUTO FORWARD, AUTO STOP or AUTO REVERSE).</p>
	<p>PERISTALTIC PUMP. If it's not included on the unit, the icon will be deactivated. If it's present, the icon will turn the peristaltic pump on/off. When you activate it, a bar will appear to let you adjust the percentage of distilled water flow. The instrument spray must also be active for correct irrigation.</p>
	<p>CRUISE CONTROL. It can be exclusively activated in ENDODONTICS mode and with AUTO FORWARD function active. This function allows the micromotor to continue turning even if you release the foot pedal lever after activating it by pushing the foot pedal lever to the right.</p> <ul style="list-style-type: none"> - Activate the cruise control function by pressing the icon; - Once the icon is pressed, an alert will pop up. To make sure that the activation of this function is desired, press YES; - The activated icon will flash red; - Move the lever to the right to activate the micromotor; - The micromotor will continue to turn even if you release the foot pedal; - To stop the micromotor while this function is activated, move the pedal lever to the right; - To deactivate this setting, press the icon again.
	<p>ATTENTION: pay particular attention to the micromotor when using it with this function active, remember that the micromotor will continue to turn if this function is not deactivated. It could be dangerous for the operator and the patient if not used with care.</p>
	<p>INSTRUMENT SPRAY. SPRAY ON: Air-water spray active when the pedal lever moves to the right. By pressing it again, you will activate the SPRAY ON/OFF function: the air-water spray will activate when you move the pedal lever to the right and press the steel lever down at the same time. By pressing it again, you will deactivate the function: SPRAY OFF.</p>
	<p>OPTICAL FIBRE. It turns on / off the LED light on the handpiece if the OPTICAL FIBRE is present on the instruments. To adjust the LED light delay, see par. 17.6.1 TIMER.</p>
	<p>MICROMOTOR ROTATION INVERSION. This function inverts the rmps rotating direction of the micromotor.</p>

20.3 SURGERY MODE

The surgery mode allows you to adjust the speed of the micromotor, the Torque value, the type of handpiece used, activate the auto reverse, auto stop and auto forward functions with related reverse time and cruise control, manage the peristaltic pump, spray, the optical fibre and inverse the micromotor rotating direction.

All these parameters can be memorized in 10 customizable and recallable programs.

By activating the pedal lever, regardless of its position, the micromotor rotates progressively from the minimum speed to the set value.



FUNCTIONS


<p>MX2 - 1</p>	<p>INSTRUMENT. It identifies which instrument you are using. If there are several identical instruments, the indicator will mark 1 / 2 / 3 according to their position on the console.</p>
<p>1:1</p>	<p>TYPE OF HANDPIECE. By pressing this icon, it is possible to choose which type of reducing or multiplying handpiece is being used on the micromotor. As the ratio changes, the number of rpms and the micromotor torque limits will vary. (For specific speed and torque values of handpieces, see par. 46)</p>
<p>PROGRAM 01</p>	<p>PROGRAM MENU. It indicates the selected program. (see par. 20.4)</p>
<p>AUTO FORWARD 5s</p>	<p>AUTO FORWARD. When you move the foot pedal lever to the right, the micromotor turns clockwise, whereas it inverts the rotating direction when it reaches the selected torque at the set time via the slide bar and then turns clockwise again. Only this function can activate the CRUISE CONTROL function.</p>
<p>AUTO REVERSE</p>	<p>AUTO REVERSE. When you move the foot pedal lever to the right, the micromotor turns clockwise, whereas it inverts the rotating direction when it reaches the set torque until you release the foot pedal.</p>
<p>AUTO STOP</p>	<p>AUTO STOP. When you move the foot pedal lever to the right, the micromotor turns clockwise and stops when it reaches the set torque. To restart the rotation, you must set the pedal lever to zero and bring it back to the right.</p>
<p>100 rpm</p>	<p>MICROMOTOR RPM. By activating the foot pedal lever, regardless of its position, the micromotor turns at the set speed. The speed can be adjusted by the cursor, the + and – keys and the hotkeys. The speed limit of the micromotor depends on the ratio of the handpiece used. (for the speed values associated with the handpiece ratio, see par. 46)</p>

	<p>TORQUE VALUE. It's a settable value ranging from 0.25Ncm to 53Ncm, depending on the handpiece used (for the torque values associated with the handpiece ratio, see par. 46). When the set resistance is reached, the micromotor will react according to the type of function activated (AUTO FORWARD, AUTO STOP or AUTO REVERSE).</p>
	<p>PERISTALTIC PUMP. If it's not included on the unit, the icon will be deactivated. If it's present, the icon will turn the peristaltic pump on/off. When you activate it, a bar will appear to let you adjust the percentage of distilled water flow. The instrument spray must also be active for correct irrigation.</p>
	<p>INSTRUMENT SPRAY. SPRAY ON: Air-water spray active when the pedal lever moves to the right. By pressing it again, you will activate the SPRAY ON/OFF function: the air-water spray will activate when you move the pedal lever to the right and press the steel lever down at the same time. By pressing it again, you will deactivate the function: SPRAY OFF.</p>
	<p>OPTICAL FIBRE. It turns on / off the LED light on the handpiece if the OPTICAL FIBRE is present on the instruments. To adjust the LED light delay, see par. 17.6.1 TIMER.</p>
	<p>MICROMOTOR ROTATION INVERSION. This function inverts the rmps rotating direction of the micromotor.</p>

20.4 PROGRAM MEMORIZATION

It is possible to save up to 10 different programs to recall specific combinations of functions and set parameters:

- Set the desired values and parameters on the screen;
- Press SAVE AS and immediately after the name of the program to which you want to associate the screen;
- By long-pressing on the name of the set program, you can change its name using the dedicated keyboard;
- With any parameter change on the screen, the program name will no longer be visible because it is no longer associated with that type of settings.




21. SCALER

When you withdraw it from its instrument housing, it is activated and regulated by the foot pedal lever.

Prophylaxis and conservative tips are used.

All variants allow you to choose between direct or proportional mode.




FUNCTIONS

<p>SCALER</p>	<p>INSTRUMENT. It identifies which instrument you are using. The number indicates the two different types of scalers. SCALER 2: when the scaler is being used without ENDO mode. SCALER 1: when you are using a Satelec scaler with ENDO or PARO mode.</p>
	<p>SCALER POWER CONTROLLER. The adjustment range goes from 0 to 100 in percentage. It can be adjusted by the cursor, the + and - keys and the shortcut keys. It can be managed in direct or proportional mode.</p>
	<p>DIRECT MODE. In direct mode, the set power of the scaler will be reached automatically when activating the foot pedal lever. If you press it again, it activates the proportional function:</p>
	<p>PROPORTIONAL MODE. In proportional mode, the power of the scaler is adjusted by the foot pedal lever: when the lever is in the zero position, the scaler is not active, whereas when you move the lever all the way to the right, the power will be progressively adjusted from a minimum to a maximum.</p>
	<p>PERISTALTIC PUMP. If it's not included on the unit, the icon will be deactivated. If it's present, the icon will turn the peristaltic pump on/off. When you activate it, a bar will appear to let you adjust the percentage of distilled water flow. The instrument spray must also be active for correct irrigation.</p>
	<p>INSTRUMENT SPRAY. SPRAY ON: Air-water spray active when the pedal lever moves to the right. By pressing it again, you will activate the SPRAY ON/OFF function: the air-water spray will activate when you move the pedal lever to the right and press the steel lever down at the same time. By pressing it again, you will deactivate the function: SPRAY OFF.</p>
	<p>OPTICAL FIBRE. It turns on / off the LED light on the handpiece if the OPTICAL FIBRE is present on the instruments. To adjust the LED light delay, see par. 17.6.1 TIMER.</p>

22. ENDO SCALER (SATELEC)

It can be used in 3 different modes, which can be selected from the drop-down menu at the top right:

- with the "ENDODONTICS" function, endodontic tips are used.
- with the "PERIODONTOLOGY" function, periodontal tips are used.
- with the "SCALING" function, tips for prophylaxis and conservative are used (for all scaler models).




FUNCTIONS

<p>SCALER</p>	<p>INSTRUMENT. It identifies which instrument you are using. The number indicates the two different types of scalers. SCALER 2: when the scaler is being used without ENDO mode. SCALER 1: when you are using a Satelec scaler with ENDO or PARO mode.</p>
	<p>SCALER POWER CONTROLLER. The adjustment range goes from 0 to 100 in percentage. It can be adjusted by the cursor, the + and - keys and the shortcut keys. It can be managed in direct or proportional mode.</p>
	<p>DIRECT MODE. In direct mode, the set power of the scaler will be reached automatically when activating the foot pedal lever. If you press it again, it activates the proportional function:</p>
	<p>PROPORTIONAL MODE. In proportional mode, the power of the scaler is adjusted by the foot pedal lever: when the lever is in the zero position, the scaler is not active, whereas when you move the lever all the way to the right, the power will be progressively adjusted from a minimum to a maximum.</p>
	<p>PERISTALTIC PUMP. If it's not included on the unit, the icon will be deactivated. If it's present, the icon will turn the peristaltic pump on/off. When you activate it, a bar will appear to let you adjust the percentage of distilled water flow. The instrument spray must also be active for correct irrigation.</p>
	<p>INSTRUMENT SPRAY. SPRAY ON: Air-water spray active when the pedal lever moves to the right. By pressing it again, you will activate the SPRAY ON/OFF function: the air-water spray will activate when you move the pedal lever to the right and press the steel lever down at the same time. By pressing it again, you will deactivate the function: SPRAY OFF.</p>
	<p>OPTICAL FIBRE. It turns on / off the LED light on the handpiece if the OPTICAL FIBRE is present on the instruments. To adjust the LED light delay, see par. 17.6.1 TIMER.</p>

23. TURBINE

When you withdraw it from its instrument housing, it is activated and regulated by the foot pedal lever. The standard base turbine works in direct mode (the set speed will be reached automatically with the activation of the pedal lever), whereas it will work in proportional mode with the proportional valve (the speed is regulated by the pedal lever).




FUNCTIONS

<p>TURBINE 1</p>	<p>INSTRUMENT. It identifies which instrument you are using. If there are several identical instruments, the indicator will mark 1 / 2 / 3 according to their position on the console.</p>
	<p>TURBINE SPEED REGULATOR. The adjustment range goes from 0 to 100 in percentage. It can be adjusted by the cursor and the + and - keys.</p>
	<p>DIRECT MODE. This icon cannot be activated. It indicates the type of turbine regulation mode. In direct mode the set turbine speed will be reached automatically when the pedal lever is activated.</p>
	<p>PROPORTIONAL MODE. This icon cannot be activated, it indicates the type of turbine regulation mode. In proportional mode, the speed of the turbine is regulated by the pedal lever: when the lever is in the zero position, the turbine is stopped; by moving it to the right up to its maximum limit, there will be progressive regulation of the speed from a minimum to a maximum.</p>
	<p>PERISTALTIC PUMP. If it's not included on the unit, the icon will be deactivated. If it's present, the icon will turn the peristaltic pump on/off. When you activate it, a bar will appear to let you adjust the percentage of distilled water flow. The instrument spray must also be active for correct irrigation.</p>
	<p>INSTRUMENT SPRAY. SPRAY ON: Air-water spray active when the pedal lever moves to the right. By pressing it again, you will activate the SPRAY ON/OFF function: the air-water spray will activate when you move the pedal lever to the right and press the steel lever down at the same time. By pressing it again, you will deactivate the function: SPRAY OFF.</p>
	<p>OPTICAL FIBRE. It turns on / off the LED light on the handpiece if the OPTICAL FIBRE is present on the instruments. To adjust the LED light delay, see par. 17.6.1 TIMER.</p>

24. SYRINGE


It is made up of a body, its hose and a metal handle complete with a removable and autoclavable spout.

It can be supplied as an option with heated water and air. The air/water supply is managed by the syringe itself.



25. CURING LIGHT

It activates automatically upon withdrawal. Read the manufacturer’s instruction for use for functions.



WARNING: Avoid directing the light towards the eyes and use protective glasses or screen.

WARNING: It is recommended to have an authorized technician check the brightness of the curing light every six months, in order to be sure that correct polymerization has been carried out.

For use, maintenance, sterilization and cleaning of the handpieces, please refer to the instructions for use contained in the packages. Miglionico S.r.l. is not held responsible for any damage to things and/or people caused by failure to comply with or omission of the aforementioned provisions.



26. HANDPIECES


INTERFACE REQUIREMENTS OF THE APPLICABLE HANDPIECES, BY DIFFERENT MANUFACTURERS										
HANDPIECE TYPE	Motive water/ cooling		Air spray		Water spray		Electric supply		Electric supply optical fibre	
	<i>l/min</i>	<i>atm</i>	<i>l/min</i>	<i>atm</i>	<i>l/min</i>	<i>atm</i>		<i>power/W</i>	<i>Vdc</i>	<i>Amp.</i>
TURBINE	35 / 50	2,2 / 4	15	3	0,15	2,5	xxxxxxx	xxxxxxx	3,5	1
MICROMOTOR	36 / 50	2,2 / 5	15	3	0,15	2,5	0 ÷ 24Vdc	65	3,5	1
SCALER	xxxxx	xxxxx	xxxxx	xxxxx	0,15	2,5	24 Vac	100	3,5	1

WARNING: For use, maintenance, sterilization and cleaning of the handpieces, please refer to the instructions for use contained in the packages. Miglionico S.r.l. is not held responsible for any damage to things and/or people caused by failure to comply with or omission of the aforementioned provisions.

The applied handpieces must comply with the 93/42 EEC directive and the CEI-EN 60601-1 and CEI EN 60601-1-2 standards, accompanied by the CE declaration of conformity, conditions and warranty certificate and use and maintenance manual.



27. SAFETY/EMERGENCY SYSTEMS



1. Backrest safety system.

2/3. Upper and lower pantograph safety system.

The safety systems listed in the image above activate when an obstacle comes between the moving parts of the unit. If activated, the entire chair lifts up to release the obstacle which hinders the movement.

When a security system is activated, an intermittent buzzer sound will go off.

WARNING: do not place hands or tools of any kind during the movement in the area indicated with the letter "B" in the image above.



HANDPIECES SAFETY SYSTEM

When the turbine, micromotor or scaler handpieces are active, the chair's movements are blocked. The use of one instrument inhibits the action of the others, except the air-water syringe which is always active.

28. HEADREST


The headrest is adjustable. To adjust its position, you must:

- Rotate the lever all the way clockwise as shown in fig. "A";
- Position the headrest as shown in fig. "B";
- Close the knob as shown in fig. "C" and make sure the headrest is blocked.

A




B



C



Simply pull the headrest to move it upwards and press the button to move it downwards to move the headrest vertically.




WARNING: When adjusting the headrest position, the patient must not place their head on it.



29. DENTIST'S CONSOLE CONFIGURATION

29.1 TOP DELIVERY VERSION




The dentist's console configuration may vary based on the options purchased.

WARNING: When using dynamic instruments, the movements of the chair are blocked; this is to avoid accidental movements of the chair while operating on the patient.



29.2 HANGING HOSES VERSION




The dentist's console configuration may vary based on the options purchased.

WARNING: When using dynamic instruments, the movements of the chair are blocked; this is to avoid accidental movements of the chair while operating on the patient.



29.3 DENTIST'S CONSOLE HANDLE REMOVAL

To remove the handle, you must push the button and simultaneously pull the handle. To place it back, you need to insert the handle into its housing and push it until the button clicks.



30. PERISTALTIC PUMP

This device is used to cool or irrigate the operating area with the desired sterile or non-sterile liquid. It can be used on the micromotor handpiece, when using handpieces with an external spray, and on the SATELEC scaler using the "steriljoint". The latter is placed between the scaling handpiece and the scaling hose and has a fitting for connecting the peristaltic pump hose.

To set up the system, proceed as follows:

- Connect the capsule complete with sterile silicone tubes to the housing (fig. A e B).
- Connect the tube with the needle to the bottle containing the liquid to be withdrawn, then connect the tube without terminal to the handpiece concerned.
- Tie the tube to the handpiece hose with the supplied ties.
- Withdraw the desired instrument from its housing.
- Press the  button.
- Press one of the 2 following spray options:  .
- Move the pedal lever to the right (the peristaltic pump will be activated automatically when the instrument is withdrawn from its housing), or to the left (only the peristaltic pump will be activated with the consequent leakage of the liquid).

To adjust the speed of the peristaltic pump and consequent variation of the flow rate, go to the settings menu.



Fig. A



Fig. B

The pump body complete with silicone tube and accessories will be supplied separately to the dental unit, in a sterile independent packaging.



31. ASSISTANT'S CONSOLE CONFIGURATION

Angled syringe with adjustable spout (the handle is removable and autoclavable at 135°).




Hoses for surgical suction

Ergonomic handle to move the console

31.1 ASSISTANT'S CONSOLE FUNCTIONS



FUNCTIONS


	Upward chair movement.
	Downward chair movement.
	Upward backrest movement.
	Downward backrest movement.
	Reset/zero position: It brings the chair to the zero position, to help the patient sit or get up from it.

	Rinsing position: It moves the backrest to the rinsing or patient rest position. A subsequent press on the icon brings the backrest back to the zero position.
	Emergency position: It brings the chair to the previously memorized Trendelenburg position.
	It pours warm water into the cup and then activates the cuspidor flushing for a set time. NOT USED
	Turns the operating light on/off. Press for 1 second. NOT USED
	It pours cold water into the cup and then activates the cuspidor flushing for a set time. NOT USED
	It activates the cuspidor flushing for a set time. NOT USED

The assistant's console configuration may vary based on the options purchased.

32. FOOT CONTROL FUNCTIONS


32.1 STANDARD FOOT CONTROL/ STANDARD WIRELESS FOOT CONTROL



	FUNCTION WHEN NO INSTRUMENT IS SELECTED	FUNCTION WHEN AN INSTRUMENT IS SELECTED
1. JOYSTICK	a – chair up b – chair down c – backrest down d – backrest up	a – chair up b – chair down c – backrest down d – backrest up
2. RESET BUTTON	It activates the RESET/ZERO POSITION. See par. 17.1	Enable / disable constant water to the instruments. See par. 17.1 Deactivates water on/off function and activates constant water to the instruments. See par. 17.1
3. SIDE LEVER (towards A)	No function.	It activates the micromotor, turbine or scaler.
3. SIDE LEVER (towards B)	Operating light ON/OFF.	Air/water spray activation from the selected instrument (if available). It stores photos when the camera is withdrawn from its housing.
4. STEEL LEVER	Water to cup activation.	Air spray activation from the selected instrument (if available). It captures photos when the camera is withdrawn from its housing.
5. PR BUTTON	It activates the PR function. See 17.1.	It activates the PR function. See 17.1.

For all specifications on the oral camera functions, please read the separate manual supplied with the camera.

32.2 PUSH FOOT CONTROL/ WIRELESS PUSH FOOT CONTROL



	FUNCTION WHEN NO INSTRUMENT IS SELECTED	FUNCTION WHEN AN INSTRUMENT IS SELECTED
--	---	---

1. JOYSTICK	a – chair up b – chair down c – backrest down d – backrest up	a – chair up b – chair down c – backrest down d – backrest up
2. RESET BUTTON	It activates the RESET/ZERO POSITION. See par. 17.1	Enable / disable constant water to the instruments. See par. 17.1 Deactivates water on/off function and activates constant water to the instruments. See par. 17.1
3. CENTRAL STEEL LEVER	No function.	It activates the micromotor, turbine or scaler.
4. SMALL STEEL LEVER	Water to cup activation.	Air spray activation from the selected instrument (if available).
5. PR BUTTON	It activates the PR function. See 17.1.	It activates the PR function. See 17.1.

For all specifications on the oral camera functions, please read the separate manual supplied with the camera.

32.3 WIRELESS FOOT CONTROL INDICATIONS

If you use the foot control approximately for 3/4 hours a day, the foot control recharge can last up to 50/60 days. You can recharge the foot control by using the power cable supplied, or via the cable to be connected to the unit. **However, it is recommended to recharge the foot control once a month to prolong the life of the battery.**

The charging condition can be monitored by the LED on the foot control and by the acoustic buzzer: here is a summary table of the various signalling methods and their meaning.


LED MEANING WHEN THE FOOT CONTROL IS CONNECTED TO THE UNIT OR TO THE POWER SUPPLY	RED LED ●	GREEN LED ●	BUZZER 🗣️
BATTERY CHARGING Foot control in standby	SLOW BLINKING	OFF	OFF
BATTERY CHARGING Foot control in use	SLOW BLINKING	1 BLINK PER SECOND	OFF
BATTERY CHARGED Foot control in standby	FIXED*	OFF	OFF
BATTERY CHARGED Foot control in use	FIXED *	1 BLINK PER SECOND	OFF

* Even if the LED does not reach the FIXED state, the battery is charged after 6 hours of charging.

LED MEANING WHEN THE FOOT CONTROL IS NOT CONNECTED TO ANY CABLE	RED LED ●	GREEN LED ●	BUZZER 🗣️
FOOT CONTROL IN USE	OFF	1 BLINK PER SECOND	OFF
BATTERY POWER > 20%	OFF	OFF	OFF
BATTERY POWER < 20%	1 BLINK EVERY 10 SECONDS	OFF	OFF
BATTERY POWER <= 10%	OFF	OFF	1 BEEPING EVERY 10 SECONDS

OPERATING SWITCH

WARNING: the switch must always be placed in the ON position.



WARNING: The battery is a lithium battery and has a 6-month warranty.



33. HOW TO CHOOSE THE INSTRUMENTS WATER SUPPLY SYSTEM

If you want to choose the water supply system for the instruments, rotate the ball valve under the dentist’s console to the desired position.



Version with independent water supply system with bottle.



34. MAINTENANCE

34.1 CLEANING AND DISINFECTION

WARNING: during maintenance, cleaning and disinfection operations, protect your eyes, respiratory tract, mouth and skin by wearing glasses with a full-face shield, masks and disposable gloves.

For the disinfection and cleansing of all surfaces (UPHOLSTERY EXCLUDES, see par. 38.2), MIGLIONICO has tested and selected Green & Clean SK produced by METASYS, therefore we recommend its exclusive use. If the customer uses any other product on the market, MIGLIONICO is not liable for warranty validity purposes.

You should not spray directly on the product’s surfaces, but rather on a soft cloth to clean all parts of the dental unit and chair.

Before disinfecting the control panels, since they are soft touch systems, it is necessary to lock the keys.

NT DENTIST CONTROL PANEL key locking procedure:

- Press this button for 3 seconds  .
- The display will be inactive for 20 seconds.



ASSISTANT CONTROL PANEL key locking procedure:

- Press these 2 icons at the same time:  
- All the LEDs flash and the control panels remains inactive for 20 seconds.

34.2 CLEANING AND CARE OF THE UPHOLSTERY

For cleaning the upholstery, Miglionico recommends using the FD 360 product from the Dürr System-Hygiene line.

Properties

FD 360 from the Dürr System-Hygiene line cleans and cares for the upholstery of the dental chair.

By using FD 360 it is possible to quickly and deeply eliminate even traces of dirt, stains or chromatic alterations that are difficult to remove because they are caused, for example, by unfixed fabric dyes. FD 360 cares and cleans thoroughly and without leaving residues; ensures a silky shiny appearance.

Product composition

FD 360 owes its action to a mixture of special surfactants, silicone compounds and protective components based on avocado oil.

Use

Use undiluted FD 360 cleaning liquid. Spray FD 360 on the surface to be cleaned and immediately wipe with a soft cloth; for deeper cleaning, apply and rub FD 360 with the special sponge. In this way, a more efficient and deeper cleaning is achieved. With the combined use of FD 360 and the special sponge included in the set, it is possible to eliminate even stubborn stains or color alterations. Subsequently, remove the excess quantity of FD 360 with a dry cloth. If necessary, use FD 360 every day.

Environmental impact

The polyethylene and polypropylene packaging can be used for both material recovery and waste-to-energy processing. For recycling purposes, rinse the bottle with water. For further information, consult the safety data sheet.

Physical data

Aspect: milky white liquid


Density: $D = 0,99 \pm 0,05 \text{ g/cm}^3$ (20 °C) pH: $3,6 \pm 0,5$



34.3 STERILIZATION


The autoclavable parts on the unit are:

DENTIST'S CONSOLE



Removable instrument protection (autoclavable at 135°)

Removable handle (autoclavable at 135°)



Tray-holder mat (weight load kg 1,5 - autoclavable at 135°)

WARNING: for cleaning and sterilization of the handpieces, please read the instructions on the packaging.



35. COMPULSORY TESTS

TO BE CARRIED OUT ANNUALLY BY AN AUTHORIZED TECHNICIAN UPON REQUEST OF THE USER

- “AF” system check
- Water and air filtering system check
- “MDS” disinfection system check
- Replace the O’-rings on the micromotor connection
- Chair safety system check
- Cuspidor safety system check
- Assistant’s arm safety system check
- Chair movements and memorizing functions check
- Internal electrical, hydric and air connections check
- Arm balancing check
- Headrest brake check
- Instrument functions check
- Instruments’ water and air pressure check
- Curing light functions check
- Amalgam separator check
- Surgical suction check
- Instruct medical and auxiliary personnel in the use and maintenance of the dental unit

DATE	DESCRIPTION	TECHNICIAN SIGNATURE

36. SAFETY TEST REPORT

Along with the dental unit, attached to this manual, you will find in paper format the result of the "TEST REPORT" carried out at our factory in Acquaviva delle Fonti (BA), relating to the "Safety Tests" carried out according to the CEI EN 60601 Standards -1 on the "NICE TOUCH" dental unit you purchased.

37. COMPATIBILITY RATES

The DM compatibility rates according to 60601-1-2 standard are:

- ESD immunity 15kV air 8kV contact
- burst 2kV/100kHz
- magnetic field: 30A/m
- CISPR Emissions 11 class A o B
- Harmonic EN 61000-3-2 class A
- immunity to RF currents in the 150kHz-80MHz range 3V modulation 80% 1kHz
- immunity to surge 1kV differential mode and 2kV common mode
- immunity to RF field:

Field (V/m)	Frequency	Modulation
3	80MHz-2700MHz	1kHz AM 80%
27	380MHz-390MHz	18Hz PM 50%
28	430MHz-470MHz	18Hz PM 50%
9	704MHz-787MHz	217Hz PM 50%
28	800MHz-960MHz	18Hz PM 50%
28	1700MHz-1990MHz	217Hz PM 50%
28	2400MHz-2570MHz	217Hz PM 50%
9	5100MHz-5800MHz	217Hz PM 50%

38. TEST REPORTS AND WARNINGS

MANUFACTURER 'S USER GUIDE AND ADVICES ABOUT ELECTROMAGNETIC FIELD (EN ISO 60601-1-2 art.6)

Aspetti di emissione		
Emission test	Compliance	Electromagnetic setting - guide
RF Emissions Cispr 11	Group 1	Miglionico Dental Unit use RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF Emissions Cispr 11	Class B	Miglionico Dental Unit is suitable for use in all establishments, including domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.
Harmonic emissions IEC 61000-3-2	Class A In compliance	It is possible to use the device in all establishments, including domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes
Voltage fluctuations/ flicker emissions IEC 61000-3-3	In compliance	

Immunity aspects			
The Dental Unit produced by Miglionico is intended for use in the electromagnetic environment specified below. The customer or the user of the device should assure that it is used in such an environment.			
Immunity test	Test level EN 60601-1-2	Compliance Level	Electromagnetic environment - guide
Electrostatic discharge (ESD) EN 61000-4-2	± 6kV contact ± 8kV air	± 6kV contact ± 8kV air	Floors should be made of wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30 %.
Burst/Fast Transient EN 61000-4-4	±2kV line power supply lines	±2kV power supply lines	Main power quality should be that of a typical commercial or hospital environment.
Surge EN 61000-4-5	±1kV differential mode	±1kV differential mode	Main power quality should be that of a typical commercial or hospital environment.
Voltage dips, short interruptions and voltage variations on power supply input lines EN 61000-4-11	< 5% UT (>95% dip in UT) fo 0,5 cycles 40% UT (60% dip in UT) for 5 cycles 70% UT (30% dip in UT) for 25 cycles < 5% UT (>95% dip in UT) for 5 seconds	< 5% UT (>95% dip in UT) fo 0,5 cycles 40% UT (60% dip in UT) for 5 cycles 70% UT (30% dip in UT) for 25 cycles < 5% UT (>95% dip in UT) for 5 seconds	Main power quality should be that of a typical commercial or hospital environment. If the user of the device requires continued operation during power mains interruptions, it is recommended that the device be powered from an uninterruptible power supply or a battery.
Power frequency magnetic field EN 61000-4-8	3 A/m	3 A/m	Magnetic power frequency fields should be that of a typical commercial or hospital environment.

Immunity aspects at r.f			
Miglionico Dental Unit is intended for use in the electromagnetic environment specified below. The customer or the user of the navigator should ensure that it is used in such an electromagnetic environment.			
Immunity test	Test level EN 60601-1-2	Compliance Level	Electromagnetic environment - guide
RF conducted EN 61000-4-6	3 Veff from 150kHz to 80MHz	3 Veff from 150kHz to 80MHz	Portable and mobile RF communications equipment should be used no closer to any part of the device, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter. Recommended separation distance: d = 1,2 √P from 150kHz to 80MHz d = 1,2 √P from 80 MHz to 800 MHz d = 2,3 √P from 800 MHz to 2,5 GHz where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in metres (m).
RF radiated EN 61000-4-3	3 Veff from 80MHz to 2,5GHz	3 Veff from 80MHz to 2,5GHz	
Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey should be less than the compliance level in each frequency range.			

Recommended separation distances between portable and mobile RF communications equipment and the device MIGLIONICO DENTAL UNIT			
Dental units by Miglionic are intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the device can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the device as recommended below, according to the maximum output power of the communications equipment.			
Rated maximum output power of transmitter (W)	Separation distance according to frequency of the transmitter (m)		
	From 150kHz to 80MHz d = 1,2 √P	Da 80MHz a 800MHz d = 1,2 √P	From 150kHz to 80MHz d = 1,2 √P
0,01	0,12	0,12	0,23
0,1	0,38	0,38	0,73
1	1,2	1,2	2,3
10	3,8	3,8	7,3
100	12	12	23
For transmitters rated at a maximum output power not listed above, the recommended separation distance d in metres (m) can be determined using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer.			
Notes:			
(1) At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.			
(1) (2) These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.			

39. MICROMOTOR RATIO VALUES

39.1 MCX MICROMOTOR

RESTORATIVE

RATIO	MAX SPEED	MIN SPEED	MAX TORQUE	LIMIT TORQUE	MIN TORQUE
20:1	2'000 RPM	5 RPM	38.0 Ncm	38.0 Ncm (100%)	3.8 Ncm (10%)
16:1	2500 RPM	6 RPM	16,4 Ncm	16,4 (100%)	1,6 (10%)
10:1	4'000 RPM	10 RPM	12.3 Ncm	12.3 Ncm (100%)	1.2 Ncm (10%)
7:1	5714 RPM	14 RPM	14 Ncm	14 Ncm (100%)	1.4 Ncm (10%)
6:1	6'666 RPM	16 RPM	12.8 Ncm	12.8 Ncm (100%)	1.3 Ncm (10%)
4:1	10'000 RPM	25 RPM	3.3 Ncm	3.3 Ncm (100%)	0.33 Ncm (10%)
2:1	20'000 RPM	50 RPM	4.2 Ncm	4.2 Ncm (100%)	0.42 Ncm (10%)
1:1	40'000 RPM	100 RPM	2.5 Ncm	2.5 Ncm (100%)	0.25 Ncm (10%)
1:2	80'000 RPM	200 RPM	1.1 Ncm	1.1 Ncm (100%)	0.11 Ncm (10%)
1:5	200'000 RPM	500 RPM	0.4 Ncm	0.4 Ncm (100%)	0.04 Ncm (10%)

ENDODONTICS

RATIO	MAX SPEED	MIN SPEED	MAX TORQUE	LIMIT TORQUE	MIN TORQUE
128	312	100	6,4	6,4	0,6
70	571	100	14	8	1,4
64	625	100	9,6	8	1
32	1250	100	12,8	8	1,3
20	2000	100	38	8	3,8
16	2000	100	16,4	8	1,6
10:1	2000	100	12,2	8	1,2
8:1	2'000	100	4.4	4.4	0.4
4:1	2'000	100	3,3	3,3	0.3
2:1	2'000	100	4.2	4.2	0.4
1:1	2'000	100	2.5	2.5	0.25

39.2 MX2 MICROMOTOR

RESTORATIVE

RATIO	MAX SPEED	MIN SPEED
20:1	2'000 RPM	5 RPM
16:1	2500 RPM	6 RPM
10:1	4'000 RPM	10 RPM
7:1	5714 RPM	14 RPM
6:1	6'666 RPM	16 RPM
4:1	10'000 RPM	25 RPM
2:1	20'000 RPM	50 RPM
1:1	40'000 RPM	100 RPM
1:2	80'000 RPM	200 RPM
1:5	200'000 RPM	500 RPM

MAX TORQUE	MIN TORQUE
53.2 Ncm (100%)	5.3 Ncm (10%)
23,0 Ncm (100%)	2,3 Ncm (10%)
17.2 Ncm (100%)	1.7 Ncm (10%)
19,6 Ncm (100%)	2,0 Ncm (10%)
17.9 Ncm (100%)	1.8 Ncm (10%)
4.6 Ncm (100%)	0.46 Ncm (10%)
5.9 Ncm (100%)	0.6 Ncm (10%)
3.5 Ncm (100%)	0.35 Ncm (10%)
1.5 Ncm (100%)	0.15 Ncm (10%)
0.6 Ncm (100%)	0.06 Ncm (10%)

ENDODONTICS

RATIO	MAX SPEED	MIN SPEED
128	312	100
70	571	100
64	625	100
32	1250	100
20	2000	100
16	2000	100
10	2000	100
8:1	2'000	100
4:1	2'000	100
2:1	2'000	100
1:1	2'000	100

MAX TORQUE	MIN TORQUE
8	0,9
8	1,9
8	1,3
8	1,8
8	5,3
8	2,3
8	1,7
6,2	0.6
4,6	0.5
6	0.6
3,5	0,35

SURGERY

RATIO	MAX SPEED	MIN SPEED
32:1	1250	3,1
20:1	2000	5
16:1	2500	6,2
1:1	40000	100
1:2	80000	200
1:5	200000	500

MAX TORQUE	MIN TORQUE
18	1,8
53,2	5,3
23	2,3
3,5	0,35
1,5	0,15
0,6	0,06

Any serious incident occurred in connection with the medical device supplied by us must be reported to the Manufacturer, the Notified Body and the Competent Authority of the Member State in which it is based.



Miglionico S.r.l.

Headquarters and production:

Via Molise, Lotti 67/68 Z.I - 70021

Acquaviva delle Fonti (BA) - ITALY

P. Iva: 05306940726

Tel +39 080 759552 - Fax +39 080 2220970

web: www.miglionico.net

CONTACTS:

Service: service@miglionico.net

Sales: export@miglionico.net

Accounting dpt: info@miglionico.net